



DENTAL UNIT WATERLINE CLEANER



ACTIVE INGREDIENTS	
Chlorhexidine Gluconate	00.12%
OTHER INGREDIENTS	99.88%
TOTAL	100.00%

LICENSED

See carton insert for additional Precautionary Statements and First Aid

PERIOD **2014-2016** LIC. NO.

8805.1

READY TO USE | DO NOT DILUTE

DO NOT PLACE MINT-A-KLEEN BOTTLE DIRECTLY ON DENTAL UNIT.

Net Contents: 16 FL OZ (473 ML)

CAUTION
KEEP OUT OF REACH OF CHILDREN



Watch our online instructional video at
www.MintAKleen.com

Questions? Please contact us directly:



109 Larrimore Lane | Danville, KY 40422 | USA
P: 866.246.2548 | F: 866.926.8246 | E: info@AnodiaSystems.com
www.AnodiaSystems.com

Dental Unit Waterline Cleaner

For the cleaning and control of microbial contamination in dental unit waterlines.
For use in any dental unit.

DIRECTIONS FOR USE:

IT IS A VIOLATION OF FEDERAL LAW TO USE THIS PRODUCT IN A MANNER INCONSISTENT WITH ITS LABELING.

INITIAL CLEANING:

Week 1: Use every night
Week 2-5: Use bi-weekly

ROUTINE CLEANING:

Use once per week. For optimal results, use at the end of the last business day of the week.

Directions for Waterline and Bottle Cleaning:

Step 1: AIR PURGE DENTAL UNIT WATERLINES

- Turn dental unit OFF. Empty water bottle and place back on dental unit.
- Turn dental unit ON.
- Air purge waterlines by activating handpiece hoses and air/water syringes until water is eliminated.

Step 2: ADD MINT-A-KLEEN

- Turn dental unit OFF. Remove water bottle, pour 4 ounces of Mint-A-Kleen into dental unit's empty water bottle, swish Mint-A-Kleen in bottle and then place water bottle on dental unit.
- Turn dental unit ON. Activate handpiece hoses and air/water syringes until Mint-A-Kleen exits.
- Turn dental unit OFF. Allow Mint-A-Kleen to treat waterlines and bottle for at least 12 hours before proceeding to Step 3.

Step 3: FLUSH DENTAL UNIT WATERLINES

- Remove water bottle, dispose of any excess Mint-A-Kleen and fill bottle with water source.
- Turn dental unit ON. Flush waterlines with water for 30 seconds to remove any excess Mint-A-Kleen.
- Continue with dental procedures.

*NOTE — For optimal cleaning, air purge and clean all instruments and tubing that contain water using the aforementioned protocol, including but not limited to: handpiece, air/water syringe, scaler, quick disconnect line.

STORAGE AND DISPOSAL

Do not contaminate water, food, or feed by storage and disposal

Storage: store in a cool dark room away from heat or open flame.

Disposal: wastes resulting from the use of this product must be disposed of on site or at an approved waste disposal facility.

Container: non-refillable container. Do not reuse or refill container. Rinse with water and place in trash or offer for recycling if available.



DENTAL UNIT WATERLINE CLEANER

For the cleaning and control of microbial contamination in dental unit waterlines. For use in any dental unit.

DIRECTIONS FOR USE:

IT IS A VIOLATION OF FEDERAL LAW TO USE THIS PRODUCT IN A MANNER INCONSISTENT WITH ITS LABELING.

INITIAL CLEANING:

Week 1: Use every night
Week 2-5: Use bi-weekly

ROUTINE CLEANING:

Use once per week. For optimal results, use at the end of the last business day of the week.

Directions for Waterline and Bottle Cleaning:

Step 1: AIR PURGE DENTAL UNIT WATERLINES

- Turn dental unit OFF. Empty water bottle and place back on dental unit.
- Turn dental unit ON.
- Air purge waterlines by activating handpiece hoses and air/water syringes until water is eliminated.

Step 2: ADD MINT-A-KLEEN

- Turn dental unit OFF. Remove water bottle, pour 4 ounces of Mint-A-Kleen into dental unit's empty water bottle, swish Mint-A-Kleen in bottle and then place water bottle on dental unit.
- Turn dental unit ON. Activate handpiece hoses and air/water syringes until Mint-A-Kleen exits.
- Turn dental unit OFF. Allow Mint-A-Kleen to treat waterlines and bottle for at least 12 hours before proceeding to Step 3.

Step 3: FLUSH DENTAL UNIT WATERLINES

- Remove water bottle, dispose of any excess Mint-A-Kleen and fill bottle with water source.
- Turn dental unit ON. Flush waterlines with water for 30 seconds to remove any excess Mint-A-Kleen.
- Continue with dental procedures.

*NOTE — For optimal cleaning, air purge and clean all instruments and tubing that contain water using the aforementioned protocol, including but not limited to: handpiece, air/water syringe, scaler, quick disconnect line.

STORAGE AND DISPOSAL

Do not contaminate water, food, or feed by storage and disposal

Storage: store in a cool dark room away from heat or open flame.

Disposal: wastes resulting from the use of this product must be disposed of on site or at an approved waste disposal facility.

Container: non-refillable container. Do not reuse or refill container. Rinse with water and place in trash or offer for recycling if available.

PRECAUTIONARY STATEMENTS:

Hazards to Humans and Domestic Animals

CAUTION: Causes moderate eye irritation. Harmful if inhaled or swallowed. Avoid contact with eyes or clothing or breathing vapor. Remove and wash contaminated clothing before reuse.

Wash thoroughly with soap and water after handling and before eating, drinking, chewing gum, using tobacco or using the toilet.

FIRST AID

IF IN EYES: Hold eye open and rinse slowly and gently with water for 15-20 minutes. Remove contact lenses, if present, after the first 5 minutes, then continue rinsing. Call a poison control center or doctor for treatment advice.

IF INHALED: Move person to fresh air. If person is not breathing, call 911 or an ambulance, then give artificial respiration, preferably mouth-to-mouth if possible. Call a poison control center or doctor for further treatment advice.

IF SWALLOWED: Call a poison control center or doctor immediately for treatment advice. Have a person sip a glass of water if able to swallow. Do not induce vomiting unless told to by a poison control center or doctor. Do not give anything to an unconscious person.

Have the product container or label with you when calling a poison control center or doctor or going for treatment.

PHYSICAL OR CHEMICAL HAZARDS

Combustible. Do not use or store near heat or open flame.

If you have any questions please contact us directly.



ANODIA
SYSTEMS

109 Larrimore Lane | Danville, KY 40422
P: 866.246.2548 | F: 866.926.8246 | E: info@AnodiaSystems.com
www.AnodiaSystems.com