

INDICATIONS FOR USE:

STERILIZATION: This solution should be used for the sterilization of heat sensitive medical equipment for which alternative methods of sterilization are not suitable. Medical equipment which should always be sterilized is that which is categorized as critical (i.e., used in procedures in which contact will be made with tissue that is normally considered sterile).

HIGH-LEVEL DISINFECTION: This solution should be used for the high-level disinfection of heat sensitive medical equipment for which sterilization is not practical. Medical equipment which should always be subjected to high-level disinfection or sterilization is that which will be used in procedures categorized as semi-critical (i.e., used in procedures in which contact will be made with mucous membranes or other body surfaces which are not normally considered to be sterile).

DIRECTIONS FOR USE:

It is a violation of Federal Law to use this product in a manner inconsistent with its labeling.

ACTIVATION: Activate the **METRICIDE® 28** solution by adding the entire contents of the *Activator Plus* activator vial which is attached to the **METRICIDE 28** solution container. Shake well. When activated, solution changes color to bluish-green. Record the date of activation on the indicated space below, in a log book or a label affixed to any secondary container used for the activated solution. See package insert for additional instructions and information regarding activated solution.

CLEANING/DECONTAMINATION: Blood and other body fluids must be thoroughly cleaned from hard, non-porous surfaces and objects before application of the disinfectant or sterilant. For complete disinfection or sterilization of medical instruments and equipment, thoroughly clean, rinse and rough dry objects before immersing in **METRICIDE 28** solution. Cleanse and rinse the lumens of hollow instruments before filling with **METRICIDE 28** solution. See package insert for additional cleaning/decontamination instructions.

STERILIZATION: Immerse medical instruments/equipment completely in **METRICIDE 28** solution for a minimum of 10 hours at 25 °C. Remove equipment from the solution using aseptic technique and rinse thoroughly with sterile water. See package insert for complete instructions/information on sterilization.

HIGH-LEVEL DISINFECTION: Immerse medical instruments/equipment completely in **METRICIDE 28** solution for a minimum of 90 minutes at 25 °C. Remove from the solution and rinse thoroughly with sterile water or with potable tap water. The quality of rinse water used is dependent upon the intended use of the instrument/equipment. See package insert for complete instructions/information on high-level disinfection.

Metricide[®]28

sterilizing &
disinfecting
solution

Metrex
Protecting People

ACTIVATED CONTENTS: ONE GALLON (128 fl. oz./3.8 Liters)

A 28-Day Glutaraldehyde,
Long-Life, Activated Dialdehyde Solution



NOTE—Contents of attached vial *must* be added to solution before this product is effective.

See *Directions for Use - Activation*.

Contains Sodium Nitrite as a corrosion inhibitor

Active ingredient:

Glutaraldehyde 2.5%

Inert ingredients: 97.5%

Total 100.0%

KEEP OUT OF REACH OF CHILDREN

DANGER: See right side panel for additional precautionary statements.

PRECAUTIONARY STATEMENTS:

Hazard to Humans and Domestic Animals

DANGER: Keep Out of Reach of Children

Direct contact is corrosive to exposed tissue, causing eye damage and skin irritation/damage. Do not get into eyes, on skin or on clothing. Wear goggles or face shield and protective gloves (butyl rubber, nitrile rubber, polyethylene or double-gloved latex), when handling or pouring. Avoid contamination of food. Use in well ventilated area in closed containers.

STATEMENT OF PRACTICAL TREATMENT:

In case of contact, immediately flush eyes or skin with plenty of water for at least 15 minutes. For eyes, get medical attention. Harmful if swallowed. Drink large quantities of water and call a physician immediately.

NOTE TO PHYSICIAN: Probable mucosal damage may contraindicate the use of gastric lavage.

STORAGE AND DISPOSAL:

Store at room temperature.

Solution Disposal: Discard residual solution in accordance with federal, state and local regulations.

Container Disposal: Do not reuse empty container. Wrap container and put in trash.

Refer to package insert for material compatibility information and more detailed usage/product data.

CONTROL NUMBER AND EXPIRATION DATE FOR UNACTIVATED SOLUTION ON BOTTLE

NOTE DATES HERE:

ACTIVATION DATE: _____

EXPIRATION DATE FOR ACTIVATED SOLUTION: _____

Do not use beyond 28 days or sooner as indicated by a 1.8% glutaraldehyde test indicator.

For product or technical information, contact Metrex at **800-841-1428** or visit our web site at **www.metrex.com**.

Manufactured By:

METREX® RESEARCH

28210 Wick Road

Romulus, Michigan 48174

MADE IN USA

Reorder No. 10-2800

70-2000-18



+H97110280022